

## EU DECLARATION OF CONFORMITY - No CETF001 rev1-1

Manufacture:

Tunstall A/S

Niels Bohrs Vej 42 DK 8660 Skanderborg

Denmark

Phone: +4587935000

This declaration is issued under the sole responsibility of the manufacture.

The manufacturer declares under our own responsibility that the following radio product(s).

| Part Number  | Description                                 |  |
|--------------|---|--|
| 2476YYWWNNNN | Fixed mounted Test Station VoIP alarm unit. |  |

Where: YY= Year (2010=10), WW=Week of production (23=week 23), NNNN= Number of produced unit this week (0001 to 9999) (WeekCount).

| Model No | Description                                | Model No | Description |  |
|----------|--|----------|-------------|--|
| K1122    | Carecom IP, fixed mounted VoIP alarm unit. |          |             |  |

Comply with the requirements of the following directive(s)

2014/30/EU Electromagnetic Compatibility Directive (EMC) 2014/35/EU Low Voltage Directive (LVD)

2011/65/EU Restriction of the use of certain hazardous substances (RoHS 2)

The conformity was checked in accordance with the following harmonised EN standard(s):

| EN 60950-1:2006+A11:2009+A1:2010+A12:2011+A2:2013 | Safety of information and communication technology equipment         |  |
|---|--|--|
| EN 55011:2009 + A1:2010                           | Industrial, scientific and medical equipment - Radio-frequency       |  |
|   | disturbance characteristics - Limits and methods of measurement      |  |
|   | CISPR 11:2009 (Modified)   |  |
| EN 55024:2010                                     | Information technology equipment - Immunity characteristics - Limits |  |
|   | and methods of measurement   |  |
|   | CISPR 24:2010  |  |
| EN 50581:2012                                     | Restriction of the use of certain hazardous substances (RoHS 2)      |  |

Signed To State 87 93 50 00

Tel::+45 87 93 50 00

Fax:+45 87 93 50 00

Date

26-11-2014

Position: I&D Manager (DK), Tunstall Healthcare Group Ltd.