

**EU DECLARATION OF CONFORMITY – N° CETF017 rev1**

**Manufacturer:** Tunstall Healthcare (UK) Ltd  
Whitley Lodge, Whitley Bridge  
Yorkshire, DN14 0HR  
UK  
Tel: +44 1977 661234

This declaration is issued under the sole responsibility of the manufacturer.  
The manufacturer declares under our own responsibility that the following radio product(s):


Product name	Product number(s)	Product Description(s)
Fast PIR	68005/89	869MHz PIR (WNX-T3) FAST

comply with the requirements of the following directive(s)

- 2014/53/EU Radio Equipment Directive (RED)**
- 2011/65/EU Restriction of the use of certain hazardous substances (RoHS)**

The conformity was checked in accordance with the following harmonised EN standard(s):

- EN 300 220-2:V3.1.1 (2017) EMC and Radio Spectrum Matters; Short Range Devices**
- EN 301 489-3:V2.1.1 (2017) EMC standard for radio equipment and services; Part 3: Specific conditions for Short-Range Devices (SRD)**
- EN 55032:2012: Electromagnetic compatibility of multimedia equipment. Emission Requirements**
- EN 50130-4: 2011 EMC Product Family Standard; Immunity Requirements for Components of Fire, Intruder and Social Alarm Systems**
- EN 60950-1: 2006/A12:2011 Safety of Information Equipment**

Signed   
\_\_\_\_\_  
Lance Kenworthy

Date 13<sup>TH</sup> MARCH 2018

Position: Group CIO & CTO, Tunstall Healthcare Group Ltd